

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052341.D
 Acq On : 15 Nov 2018 14:53
 Operator : MD\SY
 Sample : PB114828ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#05

Quant Time: Nov 16 03:52:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	326478	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	544273	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	463097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153628	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	227384	49.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.62%	
35) Dibromofluoromethane	7.59	113	177191	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
50) Toluene-d8	10.09	98	692507	50.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.30%	
62) 4-Bromofluorobenzene	12.40	95	196960	43.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.70%	

Target Compounds						Qvalue
16) Acetone	3.82	43	6388	5.03	ug/l	98
20) Methylene Chloride	4.56	84	13695	2.89	ug/l	95
43) Isopropyl Acetate	8.17	43	22423	2.94	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052341.D
Acq On : 15 Nov 2018 14:53
Operator : MD\SY
Sample : PB114828ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#05

Quant Time: Nov 16 03:52:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

