

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052343.D
 Acq On : 15 Nov 2018 15:46
 Operator : MD\SY
 Sample : PB114828ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#07

Quant Time: Nov 16 04:02:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	311504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	518373	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	449566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	152040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	217541	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
35) Dibromofluoromethane	7.59	113	165305	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	10.09	98	648901	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.40	95	189436	43.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.56%	
Target Compounds						
16) Acetone	3.82	43	6792	5.61	ug/l	99
20) Methylene Chloride	4.55	84	12491	2.73	ug/l #	88
43) Isopropyl Acetate	8.17	43	27777	3.82	ug/l #	91
80) 1,3,5-Trimethylbenzene	12.73	105	9779	1.01	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052343.D
Acq On : 15 Nov 2018 15:46
Operator : MD\SY
Sample : PB114828ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#07

Quant Time: Nov 16 04:02:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

