

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052348.D
 Acq On : 15 Nov 2018 17:59
 Operator : MD\SY
 Sample : J5767-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-111418-B

Quant Time: Nov 16 04:13:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	300942	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	510399	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	432404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	150060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	216801	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
35) Dibromofluoromethane	7.59	113	162265	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
50) Toluene-d8	10.09	98	637183	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	12.40	95	184655	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
Target Compounds						
16) Acetone	3.82	43	8947	7.65	ug/l	96
20) Methylene Chloride	4.55	84	22481	5.69	ug/l	99
43) Isopropyl Acetate	8.17	43	29634	4.14	ug/l #	89
95) Naphthalene	15.13	128	65235	14.25	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052348.D
Acq On : 15 Nov 2018 17:59
Operator : MD\SY
Sample : J5767-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
CL-01-111418-B

Quant Time: Nov 16 04:13:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

