

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052338.D
 Acq On : 15 Nov 2018 13:03
 Operator : MD\SY
 Sample : PB114828ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#02

Quant Time: Nov 16 03:35:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	320691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	538239	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	462725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	159519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	223477	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
35) Dibromofluoromethane	7.59	113	170870	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	10.09	98	681688	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
62) 4-Bromofluorobenzene	12.40	95	195480	43.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.02%	
Target Compounds						
16) Acetone	3.82	43	6925	5.55	ug/l	96
20) Methylene Chloride	4.56	84	13282	2.84	ug/l	95
43) Isopropyl Acetate	8.17	43	24715	3.27	ug/l #	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052338.D
Acq On : 15 Nov 2018 13:03
Operator : MD\SY
Sample : PB114828ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#02

Quant Time: Nov 16 03:35:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03

Instrument :

MSVOA_N

ClientSampleId :


PB114828ZHE#02

Tgt Ion:168 Resp: 320691

Ion Ratio Lower Upper

168 100

99 61.6 47.9 71.9



Abundance Ion 167.90 (167.60 to 168.60): VN052338.D (-1740) (-)

Ion 98.90 (98.60 to 99.60): VN052338.D (-1740) (-)

7.67



150000

100000

50000

0

Time-->



#16

Acetone

Concen: 5.55 ug/l

RT: 3.82 min Scan# 636

Delta R.T. 0.00 min

Lab File: VN052338.D


Acq: 15 Nov 2018 13:03

Tgt Ion: 43 Resp: 6925

Ion Ratio Lower Upper

43 100

58 32.4 24.2 36.4



Abundance Ion 43.00 (42.70 to 43.70): VN052338.D (-542) (-)

Ion 58.00 (57.70 to 58.70): VN052338.D (-542) (-)

3.82

2500

2000


1500


1000

500

0

Time-->






#20
Methylene Chloride
Concen: 2.84 ug/l
RT: 4.56 min Scan# 867
Delta R.T. 0.01 min
Lab File: VN052338.D
Acq: 15 Nov 2018 13:03

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#02

Tgt Ion:	84	Resp:	13282
Ion	Ratio	Lower	Upper
84	100		
49	149.2	118.2	177.2
51	53.1	36.1	54.1
86	57.9	51.6	77.4






Abundance Ion 83.90 (83.60 to 84.60): VN052338.D (-771) (-)

Ion 48.90 (48.60 to 49.60): VN052338.D (-771) (-)


Ion 50.90 (50.60 to 51.60): VN052338.D (-771) (-)

Ion 85.90 (85.60 to 86.60): VN052338.D (-771) (-)





#33
1,2-Dichloroethane-d4
Concen: Below ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN052338.D
Acq: 15 Nov 2018 13:03


Tgt Ion:	65	Resp:	223477
Ion	Ratio	Lower	Upper
65	100		
67	51.7	0.0	105.2



Abundance Ion 64.90 (64.60 to 65.60): VN052338.D (-1852) (-)

Ion 66.90 (66.60 to 67.60): VN052338.D (-1852) (-)





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03

Instrument :

MSVOA_N

ClientSampleId :

PB114828ZHE#02


Tgt Ion:114 Resp: 538239

Ion Ratio Lower Upper

114 100

63 24.2 0.0 45.4




88 16.1 0.0 31.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 50.00 ug/l

RT: 7.59 min Scan# 1809

Delta R.T. -0.00 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03


Tgt Ion:113 Resp: 170870

Ion Ratio Lower Upper

113 100

111 102.1 81.0 121.6



192 17.6 13.8 20.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.27 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03

Instrument :

MSVOA_N

ClientSampleId :

PB114828ZHE#02


Tgt Ion: 43 Resp: 24715

Ion Ratio Lower Upper

43 100

61 7.2 12.0 18.0#




87 7.7 8.3 12.5#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052338.D


Acq: 15 Nov 2018 13:03

Tgt Ion: 98 Resp: 681688

Ion Ratio Lower Upper



98 100


100 62.5 49.5 74.3



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03

Instrument :

MSVOA_N

ClientSampleId :

PB114828ZHE#02


Tgt Ion: 95 Resp: 195480

Ion Ratio Lower Upper

95 100

174 76.0 0.0 156.0




176 74.8 0.0 148.4



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0



#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03


Tgt Ion: 117 Resp: 462725

Ion Ratio Lower Upper

117 100

82 59.4 46.5 69.7



119 31.8 25.7 38.5




Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN052338.D

Acq: 15 Nov 2018 13:03

Instrument :

MSVOA_N

ClientSampleId :

PB114828ZHE#02


Tgt Ion:152 Resp: 159519

Ion Ratio Lower Upper

152 100

115 59.6 30.3 91.0

150 157.2 0.0 348.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

