

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111519\
 Data File : VN059271.D
 Acq On : 15 Nov 2019 17:09
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 16 02:11:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.504	0.438	13.1	70	0.00
3 P	Chloromethane	0.685	0.597	12.8	72	0.00
4 C	Vinyl Chloride	0.664	0.608	8.4#	74	0.00
5 T	Bromomethane	0.382	0.356	6.8	75	0.00
6 T	Chloroethane	0.393	0.366	6.9	76	0.00
7 T	Trichlorofluoromethane	0.794	0.784	1.3	78	0.00
8 T	Diethyl Ether	0.323	0.340	-5.3	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.479	0.478	0.2	79	0.00
10 T	Methyl Iodide	0.593	0.629	-6.1	81	0.00
11 T	Tert butyl alcohol	0.117	0.124	-6.0	84	0.00
12 CM	1,1-Dichloroethene	0.466	0.468	-0.4#	79	0.00
13 T	Acrolein	0.060	0.066	-10.0	87	0.00
14 T	Allyl chloride	0.958	0.938	2.1	78	0.00
15 T	Acrylonitrile	0.292	0.325	-11.3	84	0.00
16 T	Acetone	0.320	0.257	19.7	58	0.00
17 T	Carbon Disulfide	1.448	1.202	17.0	68	0.00
18 T	Methyl Acetate	0.759	0.848	-11.7	87	0.00
19 T	Methyl tert-butyl Ether	1.606	1.742	-8.5	84	0.00
20 T	Methylene Chloride	0.557	0.564	-1.3	81	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.498	0.6	79	0.00
22 T	Diisopropyl ether	1.810	1.910	-5.5	81	0.00
23 T	Vinyl Acetate	1.403	1.584	-12.9	81	0.00
24 P	1,1-Dichloroethane	1.011	1.032	-2.1	80	0.00
25 T	2-Butanone	0.421	0.421	0.0	74	0.00
26 T	2,2-Dichloropropane	0.884	0.861	2.6	79	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.603	-3.8	81	0.00
28 T	Bromochloromethane	0.304	0.293	3.6	77	0.00
29 T	Tetrahydrofuran	0.268	0.285	-6.3	81	0.00
30 C	Chloroform	0.970	1.018	-4.9#	83	0.00
31 T	Cyclohexane	0.957	0.874	8.7	74	0.00
32 T	1,1,1-Trichloroethane	0.815	0.852	-4.5	81	0.00
33 S	1,2-Dichloroethane-d4	0.615	0.573	6.8	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.299	0.304	-1.7	81	0.00
36 T	1,1-Dichloropropene	0.467	0.473	-1.3	79	0.00
37 T	Ethyl Acetate	0.520	0.546	-5.0	79	0.00
38 T	Carbon Tetrachloride	0.465	0.470	-1.1	79	0.00
39 T	Methylcyclohexane	0.542	0.536	1.1	75	0.00
40 TM	Benzene	1.407	1.466	-4.2	80	0.00
41 T	Methacrylonitrile	0.226	0.261	-15.5	81	0.00
42 TM	1,2-Dichloroethane	0.483	0.494	-2.3	79	0.00
43 T	Isopropyl Acetate	0.829	0.884	-6.6	82	0.00
44 TM	Trichloroethene	0.346	0.355	-2.6	80	0.00
45 C	1,2-Dichloropropane	0.382	0.407	-6.5#	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111519\
 Data File : VN059271.D
 Acq On : 15 Nov 2019 17:09
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 16 02:11:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.238	0.244	-2.5	81	0.00
47 T	Bromodichloromethane	0.476	0.516	-8.4	83	0.00
48 T	Methyl methacrylate	0.371	0.395	-6.5	80	0.00
49 T	1,4-Dioxane	0.007	0.008	-14.3	89	0.00
50 S	Toluene-d8	1.176	1.137	3.3	76	0.00
51 T	4-Methyl-2-Pentanone	0.469	0.561	-19.6	83	0.00
52 CM	Toluene	0.844	0.900	-6.6#	82	0.00
53 T	t-1,3-Dichloropropene	0.550	0.591	-7.5	83	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.639	-8.1	82	0.00
55 T	1,1,2-Trichloroethane	0.341	0.374	-9.7	86	0.00
56 T	Ethyl methacrylate	0.507	0.603	-18.9	86	0.00
57 T	1,3-Dichloropropane	0.588	0.637	-8.3	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.190	0.218	-14.7	77	0.00
59 T	2-Hexanone	0.360	0.413	-14.7	79	0.00
60 T	Dibromochloromethane	0.361	0.402	-11.4	85	0.00
61 T	1,2-Dibromoethane	0.346	0.378	-9.2	85	0.00
62 S	4-Bromofluorobenzene	0.424	0.433	-2.1	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.340	0.341	-0.3	80	0.00
65 PM	Chlorobenzene	0.975	1.055	-8.2	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.405	-10.7	86	0.00
67 C	Ethyl Benzene	1.729	1.904	-10.1#	82	0.00
68 T	m/p-Xylenes	0.648	0.713	-10.0	83	0.00
69 T	o-Xylene	0.626	0.692	-10.5	84	0.00
70 T	Styrene	0.985	1.146	-16.3	84	0.00
71 P	Bromoform	0.284	0.324	-14.1	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.637	3.844	-5.7	84	0.00
74 T	N-amyl acetate	1.541	1.710	-11.0	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.239	1.321	-6.6	87	0.00
76 T	1,2,3-Trichloropropane	1.138	1.205	-5.9	91	0.00
77 T	Bromobenzene	0.914	0.944	-3.3	85	0.00
78 T	n-propylbenzene	4.164	4.513	-8.4	83	0.00
79 T	2-Chlorotoluene	2.583	2.693	-4.3	85	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.284	-6.6	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.470	-9.0	86	0.00
82 T	4-Chlorotoluene	2.591	2.725	-5.2	83	0.00
83 T	tert-Butylbenzene	2.661	2.797	-5.1	83	0.00
84 T	1,2,4-Trimethylbenzene	2.982	3.258	-9.3	84	0.00
85 T	sec-Butylbenzene	3.461	3.775	-9.1	84	0.00
86 T	p-Isopropyltoluene	3.056	3.412	-11.6	84	0.00
87 T	1,3-Dichlorobenzene	1.565	1.679	-7.3	86	0.00
88 T	1,4-Dichlorobenzene	1.566	1.656	-5.7	86	0.00
89 T	n-Butylbenzene	2.675	2.993	-11.9	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111519\
Data File : VN059271.D
Acq On : 15 Nov 2019 17:09
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 16 02:11:13 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.614	0.639	-4.1	87	0.00
91 T	1,2-Dichlorobenzene	1.539	1.654	-7.5	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.267	4.6	87	0.00
93 T	1,2,4-Trichlorobenzene	0.825	0.983	-19.2	91	0.00
94 T	Hexachlorobutadiene	0.531	0.575	-8.3	90	0.00
95 T	Naphthalene	2.265	3.338	-47.4#	113	0.00
96 T	1,2,3-Trichlorobenzene	0.863	0.984	-14.0	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6