

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111521\
 Data File : VN069498.D
 Acq On : 15 Nov 2021 18:15
 Operator : JC/MD
 Sample : PB140771ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140771ZHE#03

Quant Time: Nov 16 00:54:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1406434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2521868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2612076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1001308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1043897	51.190	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.380%
35) Dibromofluoromethane	8.021	113	723752	47.851	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.700%
50) Toluene-d8	10.440	98	3258074	54.770	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.540%
62) 4-Bromofluorobenzene	12.731	95	1173860	53.716	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.440%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	199166	18.102	ug/l	92
20) Methylene Chloride	5.098	84	38536	2.470	ug/l	91
43) Isopropyl Acetate	8.563	43	188597	4.093	ug/l #	81
59) 2-Hexanone	11.092	43	2342755	98.277	ug/l #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111521\
Data File : VN069498.D
Acq On : 15 Nov 2021 18:15
Operator : JC/MD
Sample : PB140771ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140771ZHE#03

Quant Time: Nov 16 00:54:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

