

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
 Data File : VN075276.D
 Acq On : 15 Nov 2022 12:00
 Operator : JC\MD
 Sample : VN1115WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1115WBL01

Quant Time: Nov 16 01:01:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	149379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	273019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	96926	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	102659	51.006	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.020%
35) Dibromofluoromethane	8.171	113	83777	52.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.540%
50) Toluene-d8	10.565	98	323952	52.097	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.200%
62) 4-Bromofluorobenzene	12.847	95	99689	45.638	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
Data File : VN075276.D
Acq On : 15 Nov 2022 12:00
Operator : JC\MD
Sample : VN1115WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1115WBL01

Quant Time: Nov 16 01:01:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

