

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
 Data File : VN075279.D
 Acq On : 15 Nov 2022 13:22
 Operator : JC\MD
 Sample : N5582-05
 Misc : 1.66g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CLI-0116

Quant Time: Nov 15 14:41:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	477674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	847154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	760243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	338295	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	300160	46.637	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.280%
35) Dibromofluoromethane	8.165	113	251707	50.612	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	10.565	98	1012788	52.490	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.980%
62) 4-Bromofluorobenzene	12.847	95	335047	49.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
Data File : VN075279.D
Acq On : 15 Nov 2022 13:22
Operator : JC\MD
Sample : N5582-05
Misc : 1.66g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CLI-0116

Quant Time: Nov 15 14:41:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

