

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
 Data File : VN075282.D
 Acq On : 15 Nov 2022 14:33
 Operator : JC\MD
 Sample : N5598-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FRAC-TANK-0760533

Quant Time: Nov 16 01:02:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	161210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	289979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116001	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	108284	49.852	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.700%
35) Dibromofluoromethane	8.171	113	89190	52.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.780%
50) Toluene-d8	10.565	98	352559	53.381	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.760%
62) 4-Bromofluorobenzene	12.847	95	115545	49.803	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.600%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	11293	12.279	ug/l	100
20) Methylene Chloride	5.288	84	17804	8.682	ug/l	87
37) Ethyl Acetate	7.571	43	17613	6.094	ug/l	97
43) Isopropyl Acetate	8.688	43	80995	18.066	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
Data File : VN075282.D
Acq On : 15 Nov 2022 14:33
Operator : JC\MD
Sample : N5598-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FRAC-TANK-0760533

Quant Time: Nov 16 01:02:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

