

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
 Data File : VN075284.D
 Acq On : 15 Nov 2022 15:21
 Operator : JC\MD
 Sample : PB149039TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149039TB

Quant Time: Nov 16 01:03:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156373	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285228	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	266677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	104925	49.800	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.600%
35) Dibromofluoromethane	8.171	113	87495	52.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.500%
50) Toluene-d8	10.565	98	340693	52.444	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.880%
62) 4-Bromofluorobenzene	12.847	95	113189	49.600	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.200%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	13095	14.678	ug/l #	83
20) Methylene Chloride	5.277	84	8507	4.277	ug/l	92
37) Ethyl Acetate	7.565	43	19399	6.824	ug/l	100
43) Isopropyl Acetate	8.688	43	86751	19.672	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111522\
Data File : VN075284.D
Acq On : 15 Nov 2022 15:21
Operator : JC\MD
Sample : PB149039TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149039TB

Quant Time: Nov 16 01:03:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

