

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080104.D
 Acq On : 15 Nov 2023 12:28
 Operator : JC\MD
 Sample : VN1115WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1115WBL02

Quant Time: Nov 15 21:49:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	194852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	358326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	364493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	156771	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	162401	54.706	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.420%
35) Dibromofluoromethane	8.165	113	131819	51.977	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.960%
50) Toluene-d8	10.565	98	448165	48.003	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.000%
62) 4-Bromofluorobenzene	12.853	95	171822	55.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080104.D
Acq On : 15 Nov 2023 12:28
Operator : JC\MD
Sample : VN1115WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1115WBL02

Quant Time: Nov 15 21:49:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

