

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080107.D
 Acq On : 15 Nov 2023 13:55
 Operator : JC\MD
 Sample : PB157153TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157153TB

Quant Time: Nov 15 21:51:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	176344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	336830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	364657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	156798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	152285	56.682	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.360%
35) Dibromofluoromethane	8.171	113	118361	49.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.300%
50) Toluene-d8	10.571	98	441552	50.313	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.620%
62) 4-Bromofluorobenzene	12.853	95	172196	59.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.320%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	25280	26.788	ug/l	91
20) Methylene Chloride	5.283	84	5964	2.505	ug/l	97
37) Ethyl Acetate	7.571	43	3564	1.142	ug/l #	83
43) Isopropyl Acetate	8.694	43	141254	25.802	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080107.D
Acq On : 15 Nov 2023 13:55
Operator : JC\MD
Sample : PB157153TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157153TB

Quant Time: Nov 15 21:51:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

