

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080108.D
 Acq On : 15 Nov 2023 14:19
 Operator : JC\MD
 Sample : 05388-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Nov 15 21:52:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	187224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	355295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	384604	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	166088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156405	54.832	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.660%	
35) Dibromofluoromethane	8.165	113	121931	48.489	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.980%	
50) Toluene-d8	10.571	98	449602	48.567	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.140%	
62) 4-Bromofluorobenzene	12.853	95	177907	57.947	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.900%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	22162	22.119	ug/l	98
20) Methylene Chloride	5.277	84	8360	3.307	ug/l	95
43) Isopropyl Acetate	8.688	43	128332	22.223	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080108.D
Acq On : 15 Nov 2023 14:19
Operator : JC\MD
Sample : 05388-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Nov 15 21:52:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

