

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080109.D
 Acq On : 15 Nov 2023 14:42
 Operator : JC\MD
 Sample : 05388-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Nov 15 21:52:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	188414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	350107	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	383729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	163546	56.974	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.940%
35) Dibromofluoromethane	8.165	113	125381	50.599	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.200%
50) Toluene-d8	10.571	98	461722	50.616	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.240%
62) 4-Bromofluorobenzene	12.853	95	180297	59.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.200%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	24147	23.948	ug/l #	87
20) Methylene Chloride	5.271	84	8660	3.404	ug/l #	82
43) Isopropyl Acetate	8.694	43	132218	23.235	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080109.D
Acq On : 15 Nov 2023 14:42
Operator : JC\MD
Sample : 05388-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Nov 15 21:52:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

