

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080111.D
 Acq On : 15 Nov 2023 15:30
 Operator : JC\MD
 Sample : 05412-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-1

Quant Time: Nov 15 21:53:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	340558	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	370359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	150385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	154053	56.968	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.940%	
35) Dibromofluoromethane	8.165	113	120998	50.200	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.400%	
50) Toluene-d8	10.571	98	441222	49.725	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.440%	
62) 4-Bromofluorobenzene	12.853	95	168497	57.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.520%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	25991	27.362	ug/l	98
20) Methylene Chloride	5.289	84	9080	3.788	ug/l #	90
43) Isopropyl Acetate	8.688	43	161156	29.115	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080111.D
Acq On : 15 Nov 2023 15:30
Operator : JC\MD
Sample : 05412-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-1

Quant Time: Nov 15 21:53:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

