

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080121.D
 Acq On : 15 Nov 2023 19:28
 Operator : JC\MD
 Sample : 05432-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Nov 15 21:58:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	185417	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	359927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	375646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	159461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	160649	56.869	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.740%
35) Dibromofluoromethane	8.171	113	123467	48.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.940%
50) Toluene-d8	10.571	98	457280	48.761	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.520%
62) 4-Bromofluorobenzene	12.853	95	178463	57.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.760%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	50809	51.205	ug/l	92
20) Methylene Chloride	5.289	84	13059	5.216	ug/l	89
25) 2-Butanone	7.489	43	9212	6.340	ug/l #	72
37) Ethyl Acetate	7.571	43	3556	1.066	ug/l #	79
43) Isopropyl Acetate	8.694	43	199536	34.109	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080121.D
Acq On : 15 Nov 2023 19:28
Operator : JC\MD
Sample : 05432-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Nov 15 21:58:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

