

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080122.D
 Acq On : 15 Nov 2023 19:51
 Operator : JC\MD
 Sample : 05432-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Nov 15 21:58:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	182092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	352573	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	163631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	156917	56.562	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.120%
35) Dibromofluoromethane	8.165	113	121894	48.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	10.571	98	442379	48.156	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	12.853	95	181829	59.681	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.360%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	37465	38.446	ug/l	96
20) Methylene Chloride	5.277	84	12514	5.089	ug/l #	87
37) Ethyl Acetate	7.577	43	3380	1.034	ug/l #	74
43) Isopropyl Acetate	8.688	43	188832	32.952	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080122.D
Acq On : 15 Nov 2023 19:51
Operator : JC\MD
Sample : 05432-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Nov 15 21:58:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

