

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
 Data File : VN080123.D
 Acq On : 15 Nov 2023 20:15
 Operator : JC\MD
 Sample : 05432-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8

Quant Time: Nov 15 21:59:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/16/2023
 Supervised By : Mahesh Dadoda 11/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	188076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	361476	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	390049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	172211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	161177	56.249	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.500%			
35) Dibromofluoromethane	8.171	113	124949	48.839	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.680%			
50) Toluene-d8	10.571	98	462452	49.101	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.200%			
62) 4-Bromofluorobenzene	12.853	95	186726	59.779	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 119.560%			
Target Compounds						
						Qvalue
16) Acetone	4.436	43	42205	41.932	ug/l	91
18) Methyl Acetate	5.030	43	6249m	1.298	ug/l	
20) Methylene Chloride	5.277	84	10887	4.287	ug/l	97
43) Isopropyl Acetate	8.688	43	204737	34.848	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111523\
Data File : VN080123.D
Acq On : 15 Nov 2023 20:15
Operator : JC\MD
Sample : 05432-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8

Quant Time: Nov 15 21:59:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/16/2023
Supervised By :Mahesh Dadoda 11/17/2023

