

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111524\
 Data File : VN084873.D
 Acq On : 15 Nov 2024 16:48
 Operator : JC\MD
 Sample : P4867-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-072-IDWGW-20241113

Quant Time: Nov 16 04:14:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:38:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	25557	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	125147	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	108897	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	60280	29.339	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.800%
60) 4-Bromofluorobenzene	12.847	95	43609	23.258	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.533%
63) Toluene-d8	10.565	98	146534	28.424	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	94.733%
Target Compounds						Qvalue
37) Trichloroethene	9.353	130	2902	2.065	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111524\
Data File : VN084873.D
Acq On : 15 Nov 2024 16:48
Operator : JC\MD
Sample : P4867-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-072-IDWGW-20241113

Quant Time: Nov 16 04:14:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 01 02:38:04 2024
Response via : Initial Calibration

