

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\
 Data File : VN052353.D
 Acq On : 16 Nov 2018 9:26
 Operator : MD\SY
 Sample : VN1116WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1116WBL01

Quant Time: Nov 16 23:10:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N110118W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 02 05:39:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	42838	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	245462	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	209995	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.02	65	106562	33.52	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	111.73%	
60) 4-Bromofluorobenzene	12.40	95	85923	25.58	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	85.27%	
63) Toluene-d8	10.09	98	298254	30.20	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	100.67%	

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111618\
Data File : VN052353.D
Acq On : 16 Nov 2018 9:26
Operator : MD\SY
Sample : VN1116WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1116WBL01

Quant Time: Nov 16 23:10:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N110118W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 02 05:39:38 2018
Response via : Initial Calibration

