

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\
 Data File : VN052362.D
 Acq On : 16 Nov 2018 14:07
 Operator : MD\SY
 Sample : J5952-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 16 23:43:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	316389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	523731	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	432757	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	140227	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	222615	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
35) Dibromofluoromethane	7.59	113	169079	50.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.10%	
50) Toluene-d8	10.09	98	656074	49.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.74%	
62) 4-Bromofluorobenzene	12.40	95	184701	42.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111618\
Data File : VN052362.D
Acq On : 16 Nov 2018 14:07
Operator : MD\SY
Sample : J5952-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 16 23:43:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

