

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\  
 Data File : VN052368.D  
 Acq On : 16 Nov 2018 16:47  
 Operator : MD\SY  
 Sample : J5942-08  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 19 Sample Multiplier: 28

**Instrument :**  
 MSVOA\_N  
**ClientSampled :**  
 VW-F1

**Manual Integrations**  
**APPROVED**  
 MMDadoda  
 11/19/2018 12:46:10 PM

Quant Time: Nov 16 23:59:34 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N111418W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Nov 15 03:02:37 2018  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	298277	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	514156	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	439406	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153422	50.00	ug/l	0.00
<b>System Monitoring Compounds</b>						
33) 1,2-Dichloroethane-d4	8.03	65	219201	52.03	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.06%	
35) Dibromofluoromethane	7.59	113	169558	51.13	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.26%	
50) Toluene-d8	10.09	98	648940	49.74	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	12.40	95	190235	44.33	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.66%	
<b>Target Compounds</b>						
						<b>Qvalue</b>
16) Acetone	3.83	43	6346m	5.471	ug/l	
20) Methylene Chloride	4.56	84	132695	37.427	ug/l	98
43) Isopropyl Acetate	8.17	43	20873	2.892	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\  
 Data File : VN052368.D  
 Acq On : 16 Nov 2018 16:47  
 Operator : MD\SY  
 Sample : J5942-08  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 19 Sample Multiplier: 28

**Instrument :**  
 MSVOA\_N  
**Client Sampled :**  
 VW-F1

**Manual Integrations**  
**APPROVED**  
 MMDadoda  
 11/19/2018 12:46:10 PM

Quant Time: Nov 16 23:59:34 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N111418W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Nov 15 03:02:37 2018  
 Response via : Initial Calibration

