

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\
 Data File : VN052369.D
 Acq On : 16 Nov 2018 17:13
 Operator : MD\SY
 Sample : J5942-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VW-D1

Quant Time: Nov 17 00:01:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	305009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	512810	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	430690	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136676	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	219990	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
35) Dibromofluoromethane	7.59	113	165470	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
50) Toluene-d8	10.09	98	643006	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.40	95	181654	42.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.88%	
Target Compounds						
16) Acetone	3.82	43	7110	5.994	ug/l #	80
20) Methylene Chloride	4.55	84	14434	3.347	ug/l	96
43) Isopropyl Acetate	8.17	43	22700	3.154	ug/l #	88
95) Naphthalene	15.13	128	4791	3.543	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111618\
Data File : VN052369.D
Acq On : 16 Nov 2018 17:13
Operator : MD\SY
Sample : J5942-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VW-D1

Quant Time: Nov 17 00:01:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

