

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\
 Data File : VN052370.D
 Acq On : 16 Nov 2018 17:40
 Operator : MD\SY
 Sample : J5767-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-111418-D

Quant Time: Nov 17 00:03:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	297236	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	503123	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	419953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134941	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	215618	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
35) Dibromofluoromethane	7.60	113	161416	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
50) Toluene-d8	10.09	98	631659	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	12.40	95	176628	42.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.12%	
Target Compounds						
16) Acetone	3.82	43	14656	12.680	ug/l	92
20) Methylene Chloride	4.56	84	22908	5.897	ug/l	93
43) Isopropyl Acetate	8.17	43	20288	2.873	ug/l #	97
95) Naphthalene	15.14	128	4498	3.497	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111618\
Data File : VN052370.D
Acq On : 16 Nov 2018 17:40
Operator : MD\SY
Sample : J5767-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
CL-01-111418-D

Quant Time: Nov 17 00:03:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

