

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\
 Data File : VN052371.D
 Acq On : 16 Nov 2018 18:07
 Operator : MD\SY
 Sample : J5976-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OILY-WATER

Manual Integrations
 APPROVED

MMDadoda
 11/19/2018 12:46:17 PM

Quant Time: Nov 17 00:06:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	302707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	511090	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	423381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	215751	50.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.92%	
35) Dibromofluoromethane	7.59	113	164283	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
50) Toluene-d8	10.09	98	629166	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	
62) 4-Bromofluorobenzene	12.40	95	185452	43.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.94%	
Target Compounds						
20) Methylene Chloride	4.56	84	3741m	0.349	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111618\
Data File : VN052371.D
Acq On : 16 Nov 2018 18:07
Operator : MD\SY
Sample : J5976-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OILY-WATER

Manual Integrations
APPROVED

MMDadoda
11/19/2018 12:46:17 PM

Quant Time: Nov 17 00:06:50 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

