

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111620\
 Data File : VN064651.D
 Acq On : 16 Nov 2020 10:18
 Operator : JC/MD
 Sample : VN1116WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1116WBL01

Quant Time: Nov 17 04:02:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	234324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	475953	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	506777	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	186034	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	234083	57.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.70%	
35) Dibromofluoromethane	7.55	113	170580	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	
50) Toluene-d8	10.06	98	625195	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
62) 4-Bromofluorobenzene	12.38	95	235431	49.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111620\
Data File : VN064651.D
Acq On : 16 Nov 2020 10:18
Operator : JC/MD
Sample : VN1116WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1116WBL01

Quant Time: Nov 17 04:02:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

