

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111620\
 Data File : VN064662.D
 Acq On : 16 Nov 2020 15:24
 Operator : JC/MD
 Sample : L4774-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Nov 17 04:44:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	467621	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	490583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178582	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	227733	57.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.56%	
35) Dibromofluoromethane	7.55	113	164995	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
50) Toluene-d8	10.06	98	601472	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
62) 4-Bromofluorobenzene	12.38	95	224838	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
Target Compounds						
16) Acetone	3.78	43	5655	3.767	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111620\
Data File : VN064662.D
Acq On : 16 Nov 2020 15:24
Operator : JC/MD
Sample : L4774-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Nov 17 04:44:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

