

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069505.D
 Acq On : 16 Nov 2021 12:31
 Operator : JC/MD
 Sample : VN1116MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1116MBL01

Quant Time: Nov 16 16:34:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1376064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2395814	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2343124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	878372	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	995293	49.884	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.760%
35) Dibromofluoromethane	8.024	113	701876	48.846	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.700%
50) Toluene-d8	10.440	98	3012708	53.310	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.620%
62) 4-Bromofluorobenzene	12.731	95	1052951	50.718	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069505.D
Acq On : 16 Nov 2021 12:31
Operator : JC/MD
Sample : VN1116MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1116MBL01

Quant Time: Nov 16 16:34:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

