

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069513.D
 Acq On : 16 Nov 2021 17:16
 Operator : JC/MD
 Sample : M4641-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK-20211110

Quant Time: Nov 17 01:55:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1330303	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2329682	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2305698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	857248	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	973182	50.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.90%
35) Dibromofluoromethane	8.021	113	669723	47.93	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.86%
50) Toluene-d8	10.443	98	2961978	53.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.80%
62) 4-Bromofluorobenzene	12.731	95	1033646	51.20	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069513.D
Acq On : 16 Nov 2021 17:16
Operator : JC/MD
Sample : M4641-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK-20211110

Quant Time: Nov 17 01:55:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

