

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069515.D
 Acq On : 16 Nov 2021 18:07
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/17/2021
 Supervised By : Mahesh Dadoda 11/18/2021

Quant Time: Nov 17 00:58:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1277905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2219502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2176259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	795147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	934360	50.427	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.860%			
35) Dibromofluoromethane	8.024	113	639311	48.026	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.060%			
50) Toluene-d8	10.443	98	2810887	53.690	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.380%			
62) 4-Bromofluorobenzene	12.731	95	966460	50.250	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.500%			
Target Compounds						Qvalue
3) Chloromethane	2.309	50	3256	0.206	ug/l	98
56) Ethyl methacrylate	10.762	69	1808	2.116	ug/l #	60
59) 2-Hexanone	11.092	43	15881	8.638	ug/l	85
70) Styrene	12.291	104	3250	1.584	ug/l	96
74) N-amyl acetate	12.398	43	7939	0.296	ug/l #	94
87) 1,3-Dichlorobenzene	13.616	146	5155	0.205	ug/l	96
88) 1,4-Dichlorobenzene	13.696	146	8131m	0.330	ug/l	
89) n-Butylbenzene	13.946	91	11618	0.280	ug/l	98
91) 1,2-Dichlorobenzene	13.994	146	4867	0.204	ug/l	77
93) 1,2,4-Trichlorobenzene	15.263	180	7698	3.690	ug/l	97
95) Naphthalene	15.504	128	33796	5.323	ug/l	98
96) 1,2,3-Trichlorobenzene	15.705	180	9102	3.882	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069515.D
Acq On : 16 Nov 2021 18:07
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 17 00:58:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/17/2021
Supervised By :Mahesh Dadoda 11/18/2021

