

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069516.D
 Acq On : 16 Nov 2021 18:32
 Operator : JC/MD
 Sample : M4714-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/17/2021
 Supervised By : Mahesh Dadoda 11/18/2021

Quant Time: Nov 17 01:58:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1407076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2465455	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2455351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	915146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1034728	50.717	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.440%			
35) Dibromofluoromethane	8.024	113	711468	48.115	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.240%			
50) Toluene-d8	10.443	98	3140910	54.008	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.020%			
62) 4-Bromofluorobenzene	12.733	95	1099948	51.485	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.980%			
Target Compounds					Qvalue	
93) 1,2,4-Trichlorobenzene	15.265	180	2710	3.290	ug/l #	80
95) Naphthalene	15.509	128	11571	4.711	ug/l	95
96) 1,2,3-Trichlorobenzene	15.705	180	2938m	3.369	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069516.D
Acq On : 16 Nov 2021 18:32
Operator : JC/MD
Sample : M4714-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 17 01:58:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/17/2021
Supervised By :Mahesh Dadoda 11/18/2021

