

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069519.D
 Acq On : 16 Nov 2021 19:49
 Operator : JC/MD
 Sample : M4714-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AEI-3

Quant Time: Nov 17 02:04:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1292386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2279955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2269056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	843797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	959644	51.211	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.420%
35) Dibromofluoromethane	8.024	113	661047	48.343	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.680%
50) Toluene-d8	10.443	98	2904153	54.000	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.000%
62) 4-Bromofluorobenzene	12.731	95	1015597	51.405	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.800%
Target Compounds						
						Qvalue
30) Chloroform	7.823	83	28871	1.108	ug/l	99
44) Trichloroethene	9.220	130	8286	0.585	ug/l	100
64) Tetrachloroethene	10.980	164	3002	0.219	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069519.D
Acq On : 16 Nov 2021 19:49
Operator : JC/MD
Sample : M4714-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AEI-3

Quant Time: Nov 17 02:04:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

