

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069520.D
 Acq On : 16 Nov 2021 20:14
 Operator : JC/MD
 Sample : M4714-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AEI-4

Quant Time: Nov 17 02:07:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1375758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2438429	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2417986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	899843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1017588	51.013	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.020%
35) Dibromofluoromethane	8.024	113	706262	48.293	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.580%
50) Toluene-d8	10.443	98	3101983	53.930	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.860%
62) 4-Bromofluorobenzene	12.733	95	1082659	51.238	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.480%
Target Compounds						
						Qvalue
3) Chloromethane	2.306	50	4279	0.251	ug/l #	88
27) cis-1,2-Dichloroethene	7.327	96	6151	0.398	ug/l	44
30) Chloroform	7.820	83	40729	1.469	ug/l	99
44) Trichloroethene	9.212	130	11705	0.772	ug/l	98
64) Tetrachloroethene	10.979	164	14036	0.961	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069520.D
Acq On : 16 Nov 2021 20:14
Operator : JC/MD
Sample : M4714-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AEI-4

Quant Time: Nov 17 02:07:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

