

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069521.D
 Acq On : 16 Nov 2021 20:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 01:01:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.534	0.443	17.0	81	0.00
3 P	Chloromethane	0.619	0.490	20.8	84	0.00
4 C	Vinyl Chloride	0.620	0.506	18.4#	83	0.00
5 T	Bromomethane	0.403	0.293	27.3#	80	0.00
6 T	Chloroethane	0.419	0.325	22.4	81	0.00
7 T	Trichlorofluoromethane	0.959	0.759	20.9	80	0.00
8 T	Diethyl Ether	0.310	0.269	13.2	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.452	0.393	13.1	90	0.00
10 T	Methyl Iodide	0.554	0.506	8.7	93	0.00
11 T	Tert butyl alcohol	0.131	0.106	19.1	85	0.00
12 CM	1,1-Dichloroethene	0.405	0.355	12.3#	89	0.00
13 T	Acrolein	0.021	0.016	23.8	92	0.00
14 T	Allyl chloride	0.896	0.790	11.8	90	0.00
15 T	Acrylonitrile	0.335	0.309	7.8	94	0.00
16 T	Acetone	0.391	0.292	25.3#	76	0.00
17 T	Carbon Disulfide	0.958	0.687	28.3#	73	0.00
18 T	Methyl Acetate	1.272	1.112	12.6	92	0.00
19 T	Methyl tert-butyl Ether	1.715	1.602	6.6	95	0.00
20 T	Methylene Chloride	0.555	0.477	14.1	94	0.00
21 T	trans-1,2-Dichloroethene	0.453	0.401	11.5	92	0.00
22 T	Diisopropyl ether	1.930	1.835	4.9	96	0.00
23 T	Vinyl Acetate	1.493	1.414	5.3	93	0.00
24 P	1,1-Dichloroethane	0.983	0.898	8.6	94	0.00
25 T	2-Butanone	0.530	0.458	13.6	87	0.00
26 T	2,2-Dichloropropane	0.769	0.655	14.8	84	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.519	7.5	94	0.00
28 T	Bromochloromethane	0.503	0.508	-1.0	107	0.00
29 T	Tetrahydrofuran	0.325	0.298	8.3	90	0.00
30 C	Chloroform	1.008	0.939	6.8#	95	0.00
31 T	Cyclohexane	0.990	0.794	19.8	87	0.00
32 T	1,1,1-Trichloroethane	0.841	0.775	7.8	91	0.00
33 S	1,2-Dichloroethane-d4	0.725	0.709	2.2	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.300	0.304	-1.3	107	0.00
36 T	1,1-Dichloropropene	0.421	0.383	9.0	89	0.00
37 T	Ethyl Acetate	0.569	0.540	5.1	91	0.00
38 T	Carbon Tetrachloride	0.406	0.380	6.4	91	0.00
39 T	Methylcyclohexane	0.523	0.458	12.4	84	0.00
40 TM	Benzene	1.299	1.206	7.2	93	0.00
41 T	Methacrylonitrile	0.294	0.272	7.5	96	0.00
42 TM	1,2-Dichloroethane	0.508	0.484	4.7	95	0.00
43 T	Isopropyl Acetate	0.913	0.863	5.5	92	0.00
44 TM	Trichloroethene	0.311	0.293	5.8	94	0.00
45 C	1,2-Dichloropropane	0.347	0.337	2.9#	96	0.00
46 T	Dibromomethane	0.232	0.217	6.5	92	0.00
47 T	Bromodichloromethane	0.434	0.426	1.8	93	0.00
48 T	Methyl methacrylate	0.430	0.399	7.2	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
 Data File : VN069521.D
 Acq On : 16 Nov 2021 20:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 01:01:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	87	0.00
50 S	Toluene-d8	1.179	1.175	0.3	102	0.00
51 T	4-Methyl-2-Pentanone	0.590	0.579	1.9	92	0.00
52 CM	Toluene	0.809	0.763	5.7#	91	0.00
53 T	t-1,3-Dichloropropene	0.453	0.438	3.3	89	0.00
54 T	cis-1,3-Dichloropropene	0.495	0.483	2.4	91	0.00
55 T	1,1,2-Trichloroethane	0.326	0.322	1.2	95	0.00
56 T	Ethyl methacrylate	0.540	0.529	2.0	92	0.00
57 T	1,3-Dichloropropane	0.573	0.557	2.8	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.124	0.8	94	0.00
59 T	2-Hexanone	0.439	0.413	5.9	87	0.00
60 T	Dibromochloromethane	0.305	0.306	-0.3	91	0.00
61 T	1,2-Dibromoethane	0.330	0.326	1.2	93	0.00
62 S	4-Bromofluorobenzene	0.433	0.429	0.9	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.302	0.284	6.0	90	0.00
65 PM	Chlorobenzene	0.931	0.898	3.5	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.323	0.332	-2.8	93	0.00
67 C	Ethyl Benzene	1.716	1.659	3.3#	90	0.00
68 T	m/p-Xylenes	0.631	0.607	3.8	88	0.00
69 T	o-Xylene	0.629	0.616	2.1	90	0.00
70 T	Styrene	1.011	1.022	-1.1	91	0.00
71 P	Bromoform	0.230	0.234	-1.7	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.981	3.679	7.6	91	0.00
74 T	N-amyl acetate	1.687	1.617	4.1	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.343	1.243	7.4	92	0.00
76 T	1,2,3-Trichloropropane	1.154	1.132	1.9	96	0.00
77 T	Bromobenzene	0.910	0.854	6.2	93	0.00
78 T	n-propylbenzene	4.480	4.155	7.3	89	0.00
79 T	2-Chlorotoluene	2.774	2.533	8.7	91	0.00
80 T	1,3,5-Trimethylbenzene	3.218	2.999	6.8	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.344	0.300	12.8	83	0.00
82 T	4-Chlorotoluene	2.644	2.449	7.4	91	0.00
83 T	tert-Butylbenzene	2.820	2.719	3.6	93	0.00
84 T	1,2,4-Trimethylbenzene	3.107	2.964	4.6	90	0.00
85 T	sec-Butylbenzene	3.940	3.723	5.5	90	0.00
86 T	p-Isopropyltoluene	3.128	2.985	4.6	89	0.00
87 T	1,3-Dichlorobenzene	1.579	1.466	7.2	89	0.00
88 T	1,4-Dichlorobenzene	1.549	1.411	8.9	88	0.00
89 T	n-Butylbenzene	2.608	2.379	8.8	84	0.00
90 T	Hexachloroethane	0.513	0.493	3.9	88	0.00
91 T	1,2-Dichlorobenzene	1.502	1.460	2.8	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.217	0.214	1.4	88	0.00
93 T	1,2,4-Trichlorobenzene	0.662	0.622	6.0	82	0.00
94 T	Hexachlorobutadiene	0.390	0.368	5.6	87	0.00
95 T	Naphthalene	1.775	1.670	5.9	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111621\
Data File : VN069521.D
Acq On : 16 Nov 2021 20:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 17 01:01:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.634	0.616	2.8	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6