

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111622\
 Data File : VN075291.D
 Acq On : 16 Nov 2022 16:27
 Operator : JC\MD
 Sample : PB149082ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149082ZHE#01

Quant Time: Nov 17 04:29:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	182120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	321278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	301285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117915	48.053	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.100%
35) Dibromofluoromethane	8.171	113	99369	52.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.380%
50) Toluene-d8	10.565	98	385189	52.640	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.280%
62) 4-Bromofluorobenzene	12.847	95	128495	49.989	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.980%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	14768	14.213	ug/l	95
20) Methylene Chloride	5.295	84	10804	4.664	ug/l	84
37) Ethyl Acetate	7.565	43	21272	6.643	ug/l	100
43) Isopropyl Acetate	8.694	43	94509	19.027	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111622\
Data File : VN075291.D
Acq On : 16 Nov 2022 16:27
Operator : JC\MD
Sample : PB149082ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149082ZHE#01

Quant Time: Nov 17 04:29:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

