

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111622\
 Data File : VN075292.D
 Acq On : 16 Nov 2022 16:51
 Operator : JC\MD
 Sample : PB149082ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149082ZHE#02

Quant Time: Nov 17 04:29:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	231579	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	215562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	85817	50.457	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.920%
35) Dibromofluoromethane	8.171	113	71387	52.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.020%
50) Toluene-d8	10.571	98	275041	52.146	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.300%
62) 4-Bromofluorobenzene	12.847	95	92373	49.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.720%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	14332	19.901	ug/l	97
20) Methylene Chloride	5.277	84	10198	6.351	ug/l	87
37) Ethyl Acetate	7.565	43	21061	9.124	ug/l	99
43) Isopropyl Acetate	8.688	43	96685	27.004	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111622\
Data File : VN075292.D
Acq On : 16 Nov 2022 16:51
Operator : JC\MD
Sample : PB149082ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149082ZHE#02

Quant Time: Nov 17 04:29:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

