

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111622\
 Data File : VN075294.D
 Acq On : 16 Nov 2022 17:39
 Operator : JC\MD
 Sample : PB149082ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149082ZHE#04

Quant Time: Nov 17 04:29:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	158755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	291639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	108089	50.532	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.060%	
35) Dibromofluoromethane	8.171	113	91117	53.220	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.440%	
50) Toluene-d8	10.565	98	354943	53.436	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.880%	
62) 4-Bromofluorobenzene	12.847	95	116098	49.756	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.520%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	14710	16.241	ug/l	98
20) Methylene Chloride	5.283	84	10390	5.145	ug/l	93
37) Ethyl Acetate	7.571	43	20575	7.078	ug/l	99
43) Isopropyl Acetate	8.688	43	107030	23.738	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111622\
Data File : VN075294.D
Acq On : 16 Nov 2022 17:39
Operator : JC\MD
Sample : PB149082ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149082ZHE#04

Quant Time: Nov 17 04:29:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

