

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111623\
 Data File : VN080138.D
 Acq On : 16 Nov 2023 17:10
 Operator : JC\MD
 Sample : VN1116WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1116WBL01

Quant Time: Nov 17 01:24:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	176966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	335144	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	353003	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	155166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161352	55.220	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.440%	
35) Dibromofluoromethane	8.165	113	128146	51.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.640%	
50) Toluene-d8	10.571	98	441286	49.503	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.000%	
62) 4-Bromofluorobenzene	12.853	95	173803	51.618	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111623\
Data File : VN080138.D
Acq On : 16 Nov 2023 17:10
Operator : JC\MD
Sample : VN1116WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1116WBL01

Quant Time: Nov 17 01:24:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

