

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111720\
 Data File : VN064694.D
 Acq On : 17 Nov 2020 20:43
 Operator : JC/MD
 Sample : L4783-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7

Quant Time: Nov 18 01:51:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	241683	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	510275	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	585591	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	247235	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	238279	56.60	ug/l	0.00
Spiked Amount			Recovery	=	113.20%	
35) Dibromofluoromethane	7.55	113	175561	50.53	ug/l	0.00
Spiked Amount			Recovery	=	101.06%	
50) Toluene-d8	10.06	98	648792	48.32	ug/l	0.00
Spiked Amount			Recovery	=	96.64%	
62) 4-Bromofluorobenzene	12.37	95	298130	58.71	ug/l	0.00
Spiked Amount			Recovery	=	117.42%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.04	50	2455	0.642	ug/l #	82
13) Acrolein	3.54	56	63	0.219	ug/l #	50
16) Acetone	3.78	43	12605	7.861	ug/l	97
18) Methyl Acetate	4.28	43	3991	1.266	ug/l #	70
19) Methyl tert-butyl Ether	4.99	73	6993	0.752	ug/l	99
20) Methylene Chloride	4.51	84	979	0.308	ug/l #	65
25) 2-Butanone	6.78	43	972	0.476	ug/l #	52
29) Tetrahydrofuran	7.18	42	300	0.239	ug/l #	37
31) Cyclohexane	7.62	56	8470	1.655	ug/l #	36
36) 1,1-Dichloropropene	7.63	75	25196	4.674	ug/l #	49
37) Ethyl Acetate	6.89	43	981	0.939	ug/l #	71
38) Carbon Tetrachloride	7.63	117	29391	5.101	ug/l #	19
43) Isopropyl Acetate	8.17	43	51460	5.891	ug/l #	65
49) 1,4-Dioxane	9.17	88	96	1.653	ug/l #	1
51) 4-Methyl-2-Pentanone	10.06	43	3434	0.674	ug/l #	1
56) Ethyl methacrylate	10.41	69	254	1.880	ug/l #	72
57) 1,3-Dichloropropane	10.78	76	20390	3.100	ug/l #	45
58) 2-Chloroethyl Vinyl ether	9.87	63	31	21.548	ug/l #	51
59) 2-Hexanone	10.78	43	2369059	565.097	ug/l #	36
71) Bromoform	12.36	173	235	1.339	ug/l #	1
74) N-amyl acetate	12.03	43	234	1.238	ug/l #	21
76) 1,2,3-Trichloropropane	12.37	75	166083	27.725	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.37	75	166083	77.412	ug/l #	8
84) 1,2,4-Trimethylbenzene	13.01	105	3514	0.216	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	14.12	75	27334	22.746	ug/l #	12
95) Naphthalene	15.10	128	1082	3.697	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111720\
Data File : VN064694.D
Acq On : 17 Nov 2020 20:43
Operator : JC/MD
Sample : L4783-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7

Quant Time: Nov 18 01:51:19 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampleId :


MW-7

Tot Ion:168 Resp: 241683

Ion Ratio Lower Upper

168 100

99 69.6 53.4 80.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.63

100000

80000



60000

40000

20000

0

Time-->



#3

Chloromethane

Concen: 0.642 ug/l

RT: 2.04 min Scan# 94

Delta R.T. 0.00 min

Lab File: VN064694.D


Acq: 17 Nov 2020 20:43

Tgt Ion: 50 Resp: 2455

Ion Ratio Lower Upper

50 100

52 41.4 25.3 37.9#



Abundance Ion 49.90 (49.60 to 50.60): VN

Ion 51.90 (51.60 to 52.60): VN

2.04


1500


1000

500

0

Time-->





#13

Acrolein

Concen: 0.219 ug/l

RT: 3.54 min Scan# 561

Delta R.T. -0.02 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :


MW-7

Tot Ion: 56 Resp: 63

Ion Ratio Lower Upper


56 100

55 111.1 55.9 83.9#





Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 55.00 (54.70 to 55.70): VN0



Time-->



#16

Acetone

Concen: 7.861 ug/l

RT: 3.78 min Scan# 633

Delta R.T. -0.00 min

Lab File: VN064694.D


Acq: 17 Nov 2020 20:43

Tgt Ion: 43 Resp: 12605

Ion Ratio Lower Upper


43 100

58 28.4 21.4 32.0





Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



Time-->









#18
Methyl Acetate
Concen: 1.266 ug/l
RT: 4.28 min Scan# 791
Delta R.T. 0.00 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43




Instrument :
MSVOA_N
ClientSampled :
MW-7


Tot Ion: 43 Resp: 3991
Ion Ratio Lower Upper
43 100
74 7.2 17.0 25.4#



#19
Methyl tert-butyl Ether
Concen: 0.752 ug/l
RT: 4.99 min Scan# 1012
Delta R.T. -0.00 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43

Tgt Ion: 73 Resp: 6993
Ion Ratio Lower Upper
73 100
57 21.8 17.7 26.5









#20
Methvlene Chloride
Concen: 0.308 ug/l
RT: 4.51 min Scan# 861
Delta R.T. 0.00 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43




Instrument :
MSVOA_N
ClientSampled :
MW-7


Tot Ion: 84 Resp: 979
Ion Ratio Lower Upper
84 100
49 97.6 110.1 165.1#
51 25.3 33.6 50.4#
86 29.4 52.1 78.1#



#25
2-Butanone
Concen: 0.476 ug/l
RT: 6.78 min Scan# 1568
Delta R.T. -0.01 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43

Tgt Ion: 43 Resp: 972
Ion Ratio Lower Upper
43 100
72 0.0 18.5 27.7#









#29
Tetrahydrofuran
Concen: 0.239 ug/l
RT: 7.18 min Scan# 1692
Delta R.T. 0.01 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43




Instrument :
MSVOA_N
ClientSampled :
MW-7


Tot Ion: 42 Resp: 300
Ion Ratio Lower Upper
42 100
72 0.0 31.7 47.5#
71 0.0 29.8 44.6#



#31
Cyclohexane
Concen: 1.655 ug/l
RT: 7.62 min Scan# 1829
Delta R.T. 0.01 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43

Tgt Ion: 56 Resp: 8470
Ion Ratio Lower Upper
56 100
69 138.5 25.6 38.4#
84 67.4 63.0 94.6





#33

1,2-Dichloroethane-d4

Concen: 56.602 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. -0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :


MW-7

Tot Ion: 65 Resp: 238279

Ion Ratio Lower Upper

65 100

67 49.7 0.0 101.6



Abundance Ion 64.90 (64.60 to 65.60): VN064694.D (-1850) (-)

Ion 66.90 (66.60 to 67.60): VN064694.D (-1850) (-)

7.99

100000

80000

60000



40000

20000

0

Time-->

7.90 8.00 8.10



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2118

Delta R.T. -0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43


Tgt Ion: 114 Resp: 510275

Ion Ratio Lower Upper

114 100

63 22.7 0.0 45.6

88 17.1 0.0 32.8



Abundance Ion 113.90 (113.60 to 114.60): VN064694.D (-2025) (-)

Ion 63.00 (62.70 to 63.70): VN064694.D (-2025) (-)

Ion 87.90 (87.60 to 88.60): VN064694.D (-2025) (-)

8.55

300000


200000


100000

0

Time-->

8.50 8.60









#35
 Dibromofluoromethane
 Concen: 50.532 ug/l
 RT: 7.55 min Scan# 1807
 Delta R.T. 0.00 min
 Lab File: VN064694.D
 Acq: 17 Nov 2020 20:43




Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7


Tot Ion:113 Resp: 175561
 Ion Ratio Lower Upper
 113 100
 111 102.1 81.1 121.7
 192 16.8 12.8 19.2



#36
 1,1-Dichloropropene
 Concen: 4.674 ug/l
 RT: 7.63 min Scan# 1831
 Delta R.T. -0.13 min
 Lab File: VN064694.D
 Acq: 17 Nov 2020 20:43

Tgt Ion: 75 Resp: 25196
 Ion Ratio Lower Upper
 75 100
 110 6.2 16.0 48.0#
 77 0.0 24.0 36.0#









#37
Ethyl Acetate
Concen: 0.939 ug/l
RT: 6.89 min Scan# 1602
Delta R.T. 0.01 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43




Instrument :
MSVOA_N
ClientSampleId :
MW-7


Tot Ion: 43 Resp: 981
Ion Ratio Lower Upper
43 100
61 0.0 10.0 15.0#
70 0.0 7.1 10.7#



#38
Carbon Tetrachloride
Concen: 5.101 ug/l
RT: 7.63 min Scan# 1831
Delta R.T. -0.11 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43

Tgt Ion: 117 Resp: 29391
Ion Ratio Lower Upper
117 100
119 8.2 75.5 113.3#
121 0.0 24.4 36.6#





#43

Isopropyl Acetate

Concen: 5.891 ug/l

RT: 8.17 min Scan# 2000

Delta R.T. 0.04 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :

MW-7


Tot Ion: 43 Resp: 51460

Ion Ratio Lower Upper

43 100

61 0.6 14.0 21.0#




87 0.0 9.1 13.7#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#49

1,4-Dioxane

Concen: 1.653 ug/l

RT: 9.17 min Scan# 2312

Delta R.T. 0.01 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43


Tgt Ion: 88 Resp: 96

Ion Ratio Lower Upper

88 100

43 885.4 31.0 46.6#



58 101.0 58.4 87.6#




Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0





#50

Toluene-d8

Concen: 48.316 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :


MW-7

Tot Ion: 98 Resp: 648792




Ion Ratio Lower Upper

98 100

100 63.7 50.7 76.1



Abundance Ion 98.00 (97.70 to 98.70): VN064428.D (-2571) (-)



#51

4-Methyl-2-Pentanone

Concen: 0.674 ug/l

RT: 10.06 min Scan# 2588

Delta R.T. 0.11 min

Lab File: VN064694.D


Acq: 17 Nov 2020 20:43

Tgt Ion: 43 Resp: 3434



Ion Ratio Lower Upper


43 100

58 180.3 28.4 42.6#



Abundance Ion 43.00 (42.70 to 43.70): VN064428.D (-2535) (-)









#56
 Ethyl methacrylate
 Concen: 1.880 ug/l
 RT: 10.41 min Scan# 2697
 Delta R.T. 0.01 min
 Lab File: VN064694.D
 Acq: 17 Nov 2020 20:43




Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7


Tot Ion: 69 Resp: 254
 Ion Ratio Lower Upper
 69 100
 41 107.9 59.3 88.9#
 39 38.2 35.4 53.0



#57
 1,3-Dichloropropane
 Concen: 3.100 ug/l
 RT: 10.78 min Scan# 2812
 Delta R.T. 0.11 min
 Lab File: VN064694.D
 Acq: 17 Nov 2020 20:43

Tgt Ion: 76 Resp: 20390
 Ion Ratio Lower Upper
 76 100
 78 1.2 25.8 38.6#









#58
 2-Chloroethyl Vinyl ether
 Concen: 21.548 ug/l
 RT: 9.87 min Scan# 2527
 Delta R.T. 0.20 min
 Lab File: VN064694.D
 Acq: 17 Nov 2020 20:43




Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7


Tot Ion: 63 Resp: 31
 Ion Ratio Lower Upper
 63 100
 106 0.0 19.5 29.3#



#59
 2-Hexanone
 Concen: 565.097 ug/l
 RT: 10.78 min Scan# 2812
 Delta R.T. 0.06 min
 Lab File: VN064694.D
 Acq: 17 Nov 2020 20:43

Tgt Ion: 43 Resp: 2369059
 Ion Ratio Lower Upper
 43 100
 58 5.9 24.7 74.1#





#62

4-Bromofluorobenzene

Concen: 58.706 ug/l

RT: 12.37 min Scan# 3307

Delta R.T. -0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampleId :

MW-7


Tot Ion: 95 Resp: 298130

Ion Ratio Lower Upper

95 100

174 69.2 0.0 132.6




176 66.6 0.0 129.2



Abundance Ion 94.90 (94.60 to 95.60): VN064694.D (-3214) (-)

Ion 173.80 (173.50 to 174.50): VN064694.D (-3214) (-)

Ion 175.80 (175.50 to 176.50): VN064694.D (-3214) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2998

Delta R.T. -0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43


Tgt Ion: 117 Resp: 585591

Ion Ratio Lower Upper

117 100

82 57.7 48.1 72.1



119 31.5 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): VN064694.D (-2905) (-)

Ion 82.00 (81.70 to 82.70): VN064694.D (-2905) (-)

Ion 118.90 (118.60 to 119.60): VN064694.D (-2905) (-)





#71

Bromoform

Concen: 1.339 ug/l

RT: 12.36 min Scan# 3302

Delta R.T. 0.26 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :

MW-7


Tot Ion:173 Resp: 235

Ion Ratio Lower Upper




173 100

175 6534.0 23.7 71.1#

254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3600

Delta R.T. -0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43


Tgt Ion:152 Resp: 247235

Ion Ratio Lower Upper



152 100


115 65.0 32.0 96.0

150 157.8 0.0 353.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V









#74
N-amvl acetate
Concen: 1.238 ug/l
RT: 12.03 min Scan# 3201
Delta R.T. -0.01 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43




Instrument :
MSVOA_N
ClientSampleId :
MW-7


Tot Ion: 43 Resp: 234
Ion Ratio Lower Upper
43 100
70 54.7 29.6 44.4#
55 118.4 18.4 27.6#
61 40.6 17.8 26.6#



#76
1,2,3-Trichloropropane
Concen: 27.725 ug/l
RT: 12.37 min Scan# 3307
Delta R.T. -0.15 min
Lab File: VN064694.D
Acq: 17 Nov 2020 20:43

Tgt Ion: 75 Resp: 166083
Ion Ratio Lower Upper
75 100
77 1.0 0.0 0.0#





#81

trans-1,4-Dichloro-2-butene

Concen: 77.412 ug/l

RT: 12.37 min Scan# 3307

Delta R.T. 0.10 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :

MW-7


Tot Ion: 75 Resp: 166083

Ion Ratio Lower Upper

75 100

53 0.4 84.0 126.0#




89 0.0 36.3 54.5#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#84

1,2,4-Trimethylbenzene

Concen: 0.216 ug/l

RT: 13.01 min Scan# 3505

Delta R.T. -0.00 min

Lab File: VN064694.D


Acq: 17 Nov 2020 20:43

Tgt Ion: 105 Resp: 3514

Ion Ratio Lower Upper



105 100


120 45.6 21.9 65.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#92

1,2-Dibromo-3-Chloropropane

Concen: 22.746 ug/l

RT: 14.12 min Scan# 3849

Delta R.T. -0.13 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43

Instrument :

MSVOA_N

ClientSampled :

MW-7


Tot Ion: 75 Resp: 27334

Ion Ratio Lower Upper

75 100

155 1.7 34.6 103.9#




157 1.6 43.0 128.8#



Abundance Ion 74.90 (74.60 to 75.60): VN064428.D (-3874) (-)

Ion 154.80 (154.50 to 155.50): VN064428.D (-3874) (-)

Ion 156.80 (156.50 to 157.50): VN064428.D (-3874) (-)



#95

Naphthalene

Concen: 3.697 ug/l

RT: 15.10 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN064694.D

Acq: 17 Nov 2020 20:43


Tgt Ion:128 Resp: 1082

Ion Ratio Lower Upper

128 100

127 11.3 10.5 15.7

129 6.0 8.8 13.2#



Abundance Ion 128.00 (127.70 to 128.70): VN064428.D (-4139) (-)

Ion 127.00 (126.70 to 127.70): VN064428.D (-4139) (-)

Ion 129.00 (128.70 to 129.70): VN064428.D (-4139) (-)

