

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111720\
 Data File : VN064676.D
 Acq On : 17 Nov 2020 13:30
 Operator : JC/MD
 Sample : L4772-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-111620

Quant Time: Nov 18 02:17:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	219350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	435865	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	454823	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	159781	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	214366	56.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.22%	
35) Dibromofluoromethane	7.55	113	158212	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
50) Toluene-d8	10.06	98	579517	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.37	95	205929	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
Target Compounds						
3) Chloromethane	2.04	50	2980	0.859	ug/l	97
16) Acetone	3.78	43	6020	4.136	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111720\
Data File : VN064676.D
Acq On : 17 Nov 2020 13:30
Operator : JC/MD
Sample : L4772-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-111620

Quant Time: Nov 18 02:17:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

