

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111720\
 Data File : VN064680.D
 Acq On : 17 Nov 2020 15:07
 Operator : JC/MD
 Sample : L4772-05DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3RDL

Manual Integrations
 APPROVED

MMDadoda
 11/18/2020 5:15:24 PM

Quant Time: Nov 18 02:41:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	231464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	462399	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	500495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	231324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	221807	55.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.04%	
35) Dibromofluoromethane	7.55	113	163043	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
50) Toluene-d8	10.06	98	609365	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	12.38	95	261394	56.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.60%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.05	83	5486m	1.253	ug/l	
67) Ethyl Benzene	11.48	91	102308	5.242	ug/l	95
68) m/p-Xylenes	11.60	106	88071	12.443	ug/l	97
69) o-Xylene	11.92	106	10302	1.506	ug/l	92
73) Isopropylbenzene	12.22	105	297882	16.201	ug/l	99
78) n-propylbenzene	12.56	91	558229	26.631	ug/l	98
80) 1,3,5-Trimethylbenzene	12.71	105	119821	7.899	ug/l	99
83) tert-Butylbenzene	12.97	119	30936	2.611	ug/l	98
84) 1,2,4-Trimethylbenzene	13.01	105	2259015	148.149	ug/l	99
85) sec-Butylbenzene	13.14	105	160045	10.667	ug/l #	75
86) p-Isopropyltoluene	13.26	119	75014	5.559	ug/l	98
89) n-Butylbenzene	13.59	91	91144	7.855	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111720\
Data File : VN064680.D
Acq On : 17 Nov 2020 15:07
Operator : JC/MD
Sample : L4772-05DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3RDL

Manual Integrations
APPROVED

MMDadoda
11/18/2020 5:15:24 PM

Quant Time: Nov 18 02:41:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

