

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111720\
 Data File : VN064682.D
 Acq On : 17 Nov 2020 15:55
 Operator : JC/MD
 Sample : L4772-08DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FD-111620DL

Manual Integrations
 APPROVED

MMDadoda
 11/18/2020 5:15:47 PM

Quant Time: Nov 18 02:47:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	235903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	472096	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	508063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	231277	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	226725	55.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.36%	
35) Dibromofluoromethane	7.55	113	164127	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
50) Toluene-d8	10.06	98	616949	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	12.38	95	267793	57.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.00%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.73	59	3014m	6.651	ug/l	
39) Methylcyclohexane	9.04	83	5210	1.165	ug/l #	81
67) Ethyl Benzene	11.48	91	107208	5.411	ug/l	94
68) m/p-Xylenes	11.60	106	86968	12.104	ug/l	100
69) o-Xylene	11.92	106	10244	1.475	ug/l	80
73) Isopropylbenzene	12.22	105	307914	16.750	ug/l	100
78) n-propylbenzene	12.57	91	572362	27.310	ug/l	100
80) 1,3,5-Trimethylbenzene	12.70	105	122659	8.088	ug/l	96
83) tert-Butylbenzene	12.96	119	31432	2.654	ug/l	96
84) 1,2,4-Trimethylbenzene	13.01	105	2291993	150.343	ug/l	100
85) sec-Butylbenzene	13.14	105	162419	10.828	ug/l #	77
86) p-Isopropyltoluene	13.26	119	75301	5.581	ug/l	99
89) n-Butylbenzene	13.59	91	91460	7.884	ug/l #	78
95) Naphthalene	15.11	128	6644	4.109	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111720\
Data File : VN064682.D
Acq On : 17 Nov 2020 15:55
Operator : JC/MD
Sample : L4772-08DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FD-111620DL

Manual Integrations
APPROVED

MMDadoda
11/18/2020 5:15:47 PM

Quant Time: Nov 18 02:47:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

