


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111720\
Data File : VN064671.D
Acq On : 17 Nov 2020 10:04
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Title : SW846 8260
Last Update : Wed Nov 04 12:09:03 2020



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	8875	PASS
75	95	30	60	56.9	22141	PASS
95	95	100	100	100.0	38914	PASS
96	95	5	9	6.2	2415	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	69.1	26898	PASS
175	174	5	9	8.5	2295	PASS
176	174	95	101	96.9	26066	PASS
177	176	5	9	5.8	1521	PASS