

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111721\
 Data File : VN069536.D
 Acq On : 17 Nov 2021 17:53
 Operator : JC/MD
 Sample : PB140850TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140850TB

Quant Time: Nov 18 07:12:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1230341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2172877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2223797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	813961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	920312	51.589	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.180%
35) Dibromofluoromethane	8.024	113	623844	47.870	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.740%
50) Toluene-d8	10.443	98	2812194	54.867	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.740%
62) 4-Bromofluorobenzene	12.734	95	987446	52.443	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.880%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	89336	9.282	ug/l	96
20) Methylene Chloride	5.093	84	48602	3.561	ug/l	92
25) 2-Butanone	7.345	43	65507	5.026	ug/l	91
43) Isopropyl Acetate	8.566	43	137626	3.467	ug/l #	87
59) 2-Hexanone	11.092	43	2078807	100.973	ug/l #	49

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111721\
Data File : VN069536.D
Acq On : 17 Nov 2021 17:53
Operator : JC/MD
Sample : PB140850TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140850TB

Quant Time: Nov 18 07:12:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

