

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111721\
 Data File : VN069538.D
 Acq On : 17 Nov 2021 18:45
 Operator : JC/MD
 Sample : M4695-01
 Misc : 5.09g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 218

Quant Time: Nov 18 07:12:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1321870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2325850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2218801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	951870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	996260	51.979	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.960%
35) Dibromofluoromethane	8.019	113	642196	46.037	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.080%
50) Toluene-d8	10.441	98	2935176	53.500	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.000%
62) 4-Bromofluorobenzene	12.734	95	1021009	50.659	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.320%
Target Compounds						
95) Naphthalene	15.507	128	7912	4.622	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111721\
Data File : VN069538.D
Acq On : 17 Nov 2021 18:45
Operator : JC/MD
Sample : M4695-01
Misc : 5.09g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
218

Quant Time: Nov 18 07:12:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

