

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111722\
 Data File : VN075299.D
 Acq On : 17 Nov 2022 11:37
 Operator : JC\MD
 Sample : VN1117MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1117MBL01

Quant Time: Nov 18 00:59:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	165485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	288625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107535	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	102678	46.050	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.100%
35) Dibromofluoromethane	8.171	113	86260	50.910	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.820%
50) Toluene-d8	10.565	98	330857	50.330	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.660%
62) 4-Bromofluorobenzene	12.847	95	104516	45.260	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111722\
Data File : VN075299.D
Acq On : 17 Nov 2022 11:37
Operator : JC\MD
Sample : VN1117MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1117MBL01

Quant Time: Nov 18 00:59:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

