

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111722\
 Data File : VN075300.D
 Acq On : 17 Nov 2022 12:01
 Operator : JC\MD
 Sample : VN1117WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1117WBL01

Quant Time: Nov 18 00:59:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	156589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	281783	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103063	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	103127	48.879	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.760%
35) Dibromofluoromethane	8.171	113	86968	52.574	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.140%
50) Toluene-d8	10.565	98	323691	50.436	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.847	95	103925	46.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111722\
Data File : VN075300.D
Acq On : 17 Nov 2022 12:01
Operator : JC\MD
Sample : VN1117WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1117WBL01

Quant Time: Nov 18 00:59:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

