

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111722\
 Data File : VN075303.D
 Acq On : 17 Nov 2022 13:34
 Operator : JC\MD
 Sample : N5675-07 200X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1913

Quant Time: Nov 18 01:00:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	158285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	282277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	103913	48.724	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.440%
35) Dibromofluoromethane	8.171	113	86022	51.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.820%
50) Toluene-d8	10.565	98	325196	50.582	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.847	95	104814	46.410	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
16) Acetone	4.453	43	3715	4.114	ug/l	Qvalue # 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111722\
Data File : VN075303.D
Acq On : 17 Nov 2022 13:34
Operator : JC\MD
Sample : N5675-07 200X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1913

Quant Time: Nov 18 01:00:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

