

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111723\
 Data File : VN080169.D
 Acq On : 17 Nov 2023 17:40
 Operator : JC\MD
 Sample : 05462-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-21

Quant Time: Nov 18 02:34:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	167009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	321551	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	358061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	156605	56.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.580%	
35) Dibromofluoromethane	8.171	113	120965	50.984	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.960%	
50) Toluene-d8	10.571	98	444293	51.947	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.900%	
62) 4-Bromofluorobenzene	12.853	95	183256	56.726	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	42053	45.404	ug/l	95
20) Methylene Chloride	5.277	84	59895	30.880	ug/l	94
40) Benzene	8.612	78	248721	25.476	ug/l	100
43) Isopropyl Acetate	8.694	43	129370	27.901	ug/l #	84
52) Toluene	10.635	92	162940	26.668	ug/l	98
67) Ethyl Benzene	11.965	91	32944	2.282	ug/l	93
68) m/p-Xylenes	12.071	106	45869	8.714	ug/l	92
69) o-Xylene	12.400	106	23249	4.686	ug/l	97
84) 1,2,4-Trimethylbenzene	13.488	105	11246	1.010	ug/l	84
95) Naphthalene	15.641	128	70353	7.385	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111723\
Data File : VN080169.D
Acq On : 17 Nov 2023 17:40
Operator : JC\MD
Sample : 05462-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-21

Quant Time: Nov 18 02:34:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

