

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111723\
 Data File : VN080176.D
 Acq On : 17 Nov 2023 20:27
 Operator : JC\MD
 Sample : 05449-31
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SED-03-G

Quant Time: Nov 18 02:38:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	164662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	322172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	340569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	150062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	149753	55.080	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.160%		
35) Dibromofluoromethane	8.171	113	117821	49.563	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.120%		
50) Toluene-d8	10.571	98	425729	49.680	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.360%		
62) 4-Bromofluorobenzene	12.853	95	170191	52.580	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.160%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	35408	38.774	ug/l #	86
18) Methyl Acetate	5.041	43	3256	1.110	ug/l #	60
20) Methylene Chloride	5.277	84	8264	4.321	ug/l #	90
25) 2-Butanone	7.500	43	6499	5.990	ug/l	100
43) Isopropyl Acetate	8.694	43	132212	28.459	ug/l #	82
95) Naphthalene	15.641	128	8569	1.002	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111723\
Data File : VN080176.D
Acq On : 17 Nov 2023 20:27
Operator : JC\MD
Sample : 05449-31
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SED-03-G

Quant Time: Nov 18 02:38:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

